

on infrared absorption spectroscopy and is identified as follows:

RFCA-0915-228, "Environnement S.A. Model CO12e Carbon Monoxide Analyzer", an infrared absorption spectroscopy technique operated on a full scale range of 0–50 ppm, at any temperature in the range of 10 °C to 35 °C, with a teflon sample particulate filter with the following software settings: Automatic response time ON; Automatic "ZERO-REF" cycle either ON or OFF and with or without the following options: ESTEL Analog Input/Output Board, LCD color touch screen and Carbon Dioxide CO₂ sensor.

This application for a reference method determination for this CO method was received by the Office of Research and Development on July 20, 2015. This analyzer is commercially available from the applicant, Environnement S.A., 111, Boulevard Robespierre, 78300 Poissy France.

The new equivalent method for O₃ is an automated method that utilizes a measurement principle based on non-dispersive ultraviolet absorption photometry. The newly designated equivalent method for O₃ is identified as follows:

EQOA-1015-229, "Teledyne Advanced Pollution Instrumentation, Model 430 Ozone Analyzer", operated with a full scale range between 0–500 ppb, at any operating temperature from 5 °C to 40 °C, with a sample particulate filter, with a 100–240V AC to DC power adapter or a 12V DC source capable of providing 9 watts of power, in accordance with the associated instrument manual, and with or without any of the following options: Internal long-life pump, external long-life pump, external portable battery pack, external communication and data monitoring interfaces.

The application for an equivalent method determination for this candidate method was received by the Office of Research and Development on August 27, 2015. The analyzer is commercially available from the applicant, Teledyne Advanced Pollution Instrumentation, Inc., 9480 Carroll Park Drive, San Diego, CA 92121-2251.

Representative test analyzers have been tested in accordance with the applicable test procedures specified in 40 CFR part 53, as amended on August 31, 2011. After reviewing the results of those tests and other information submitted by the applicant, EPA has determined, in accordance with part 53, that these methods should be designated as a reference or equivalent method.

As a designated reference or equivalent method, these methods are acceptable for use by states and other air monitoring agencies under the requirements of 40 CFR part 58, Ambient Air Quality Surveillance. For such purposes, each method must be

used in strict accordance with the operation or instruction manual associated with the method and subject to any specifications and limitations (e.g., configuration or operational settings) specified in the designated method description (see the identification of the method above).

Use of the method also should be in general accordance with the guidance and recommendations of applicable sections of the "Quality Assurance Handbook for Air Pollution Measurement Systems, Volume I," EPA/600/R-94/038a and "Quality Assurance Handbook for Air Pollution Measurement Systems, Volume II, Ambient Air Quality Monitoring Program," EPA-454/B-13-003, (both available at <http://www.epa.gov/ttn/amtic/qalist.html>). Provisions concerning modification of such methods by users are specified under Section 2.8 (Modifications of Methods by Users) of Appendix C to 40 CFR part 58.

Consistent or repeated noncompliance with any of these conditions should be reported to: Director, Human Exposure and Atmospheric Sciences Division (MD-E205-01), National Exposure Research Laboratory, U.S. Environmental Protection Agency, Research Triangle Park, North Carolina 27711.

Designation of these reference and equivalent methods is intended to assist the States in establishing and operating their air quality surveillance systems under 40 CFR part 58. Questions concerning the commercial availability or technical aspects of the method should be directed to the applicant.

Dated: November 6, 2015.

Jennifer Orme-Zavaleta,
Director, National Exposure Research Laboratory.

[FR Doc. 2015-29492 Filed 11-18-15; 8:45 am]

BILLING CODE 6560-50-P

FEDERAL DEPOSIT INSURANCE CORPORATION

Notice of Termination; 10454 The Royal Palm Bank of Florida, Naples, FL

The Federal Deposit Insurance Corporation (FDIC), as Receiver for 10454 The Royal Palm Bank of Florida, Naples, FL (Receiver) has been authorized to take all actions necessary to terminate the receivership estate of The Royal Palm Bank of Florida (Receivership Estate); The Receiver has made all dividend distributions required by law.

The Receiver has further irrevocably authorized and appointed FDIC-

Corporate as its attorney-in-fact to execute and file any and all documents that may be required to be executed by the Receiver which FDIC-Corporate, in its sole discretion, deems necessary; including but not limited to releases, discharges, satisfactions, endorsements, assignments and deeds.

Effective November 1, 2015 the Receivership Estate has been terminated, the Receiver discharged, and the Receivership Estate has ceased to exist as a legal entity.

Dated: November 16, 2015.

Federal Deposit Insurance Corporation.

Robert E. Feldman,
Executive Secretary.

[FR Doc. 2015-29541 Filed 11-18-15; 8:45 am]

BILLING CODE 6714-01-P

FEDERAL DEPOSIT INSURANCE CORPORATION

Notice of Termination, 10472 Gold Canyon Bank, Gold Canyon, Arizona

The Federal Deposit Insurance Corporation (FDIC), as Receiver for 10472 Gild Canyon Bank, Gold Canyon, Arizona (Receiver) has been authorized to take all actions necessary to terminate the receivership estate of Gold Canyon Bank (Receivership Estate); The Receiver has made all dividend distributions required by law.

The Receiver has further irrevocably authorized and appointed FDIC-Corporate as its attorney-in-fact to execute and file any and all documents that may be required to be executed by the Receiver which FDIC-Corporate, in its sole discretion, deems necessary; including but not limited to releases, discharges, satisfactions, endorsements, assignments and deeds.

Effective November 01, 2015 the Receivership Estate has been terminated, the Receiver discharged, and the Receivership Estate has ceased to exist as a legal entity.

Dated: November 16, 2015.

Federal Deposit Insurance Corporation.

Robert E. Feldman,
Executive Secretary.

[FR Doc. 2015-29526 Filed 11-18-15; 8:45 am]

BILLING CODE 6714-01-P

FEDERAL MINE SAFETY AND HEALTH REVIEW COMMISSION

Sunshine Act Meeting

November 17, 2015.

TIME AND DATE: 10:00 a.m., Thursday, December 3, 2015.